



**thsti**

ट्रान्सलेशनल स्वास्थ्य विज्ञान  
एवं प्रौद्योगिकी संस्थान

TRANSLATIONAL HEALTH SCIENCE  
AND TECHNOLOGY INSTITUTE

**3<sup>rd</sup> Mile Stone, Faridabad-Gurgaon Expressway, Faridabad – 121001**  
**Phone: +91-129-2876432**

**TENDER NO - THSTI/NIT/05/2018-19**  
**Automatic Digital Polarimeter**

**Corrigendum No. – 1**

<b>SL No.</b>	<b>Tender Part</b>	<b>Existing Entry in Tender (FOR)</b>	<b>Amended Entry in Tender (TO BE READ AS)</b>
<b>1</b>	Technical Specification	Sample Tubes: cylindrical glass cell	<b>Sample Tubes: cylindrical glass cell/ Temptrol Cell/ Stainless Steel Cell</b>
<b>2.</b>	Technical Specification	Temperature control probe with Temperature range: 15-40 <sup>o</sup> C (accuracy +/- 0.1 Deg C)	<b>Temperature control probe with Temperature range: 15-35<sup>o</sup> C (accuracy +/- 0.2 Deg C)</b>

**The last date for submission of online bids duly completed in all respects including the above amendment is same as given (28/06/2018 up to 1500 hrs.)**

All other terms and conditions of the subject tender enquiry will remain same.

**Section Officer (S&P)**  
**THSTI**