

**TRANSLATIONAL HEALTH SCIENCE & TECHNOLOGY INSTITUTE**

DATE: 07.10.2019

**CORRIGENDUM**

**TENDER REFERENCE NO. THSTI/NIT/13/19-20 DATED 04 OCT 2019**


TENDER FOR PROVIDING BENCH TOP MULTIPURPOSE CENTRIFUGE (REFRIGERATED) QTY-03  
TO TRANSLATION HEALTH SCIENCE & TECHNOLOGY INSTITUTE, FARIDABAD.

S.NO.	REF CLAUSE/POINT/SUBSECTION	ADDENDUM/AMENDMENT TO THE BID DOCUMENT
01	POINT NO. 11	Noise level should be equal to or less than 64db

Bidders are instructed to consider the point 11 as mentioned above with regards to the published specification and submit the bid to Thsti.

This Addendum forms the part of the Tender Document (Techno commercial Bid) and is required to be signed and sealed by the bidder while submitting bids to THSTI.

**\*All other conditions of the contract would remain the same. \***

  
( SATISH KUMAR  
SECTION OFFICER (S&P)

DATE: 07.10.2019